

# Ventyv<sup>®</sup>

## Nitrile PF Exam

PLUS 5.0 Black



Ventyv<sup>®</sup>  
Outsmart Infection<sup>®</sup>



sritrangusa.com | ventyv.com | Hello@sritrangusa.com | 1.888.483.6898 |  
+1 (813) 606-4301 | 5820 West Cypress Street, Suite H | Tampa, FL 33607  
Issued: 20210913

# Ventyv<sup>®</sup>

Nitrile PF Exam  
PLUS 5.0 Black



Ventyv<sup>®</sup> Nitrile Powder-Free PLUS 5.0 Black (Bull) Exam Glove has the fit and feel of a traditional nitrile glove and is ideal for use where extra protection is needed in clinical settings.



## Bull



Reorder #	Size	Gloves/ Box <sup>1</sup>	Boxes/ Case
10318108	XS	100	10
10319108	S	100	10
10320108	M	100	10
10321108	L	100	10
10322108	XL	100	10
10323108	XXL	100	10

### PRODUCT ATTRIBUTES

- Black
- Nitrile<sup>2</sup>
- Powder-free
- Polymer Coated
- Tested for Use with Chemotherapy Drugs<sup>4</sup>
- Tested for Use with Fentanyl<sup>4</sup>
- Textured

### PHYSICAL ATTRIBUTES<sup>3</sup>

- Length: 220 mm (XS, S)  
230 mm (M, L, XL)
- Width: 70 ± 10 mm (XS)  
80 ± 10 mm (S)  
95 ± 10 mm (M)  
110 ± 10 mm (L)  
120 ± 10 mm (XL)  
130 ± 10 mm (XXL)

### THICKNESS (Single Wall)

- Finger: 0.13 mm / 5 mils
- Palm: 0.10 mm / 4 mils

### PHYSICAL PROPERTIES

- Tensile strength before aging: 44 MPa
- Tensile strength after aging: 39 MPa
- Elongation before aging: 540%
- Elongation after aging: 481%



**80%** Inner boxes made with 80% recycled material

**90%** Outer cartons made with 90% recycled material

1. By weight
2. This product is not made with Natural Rubber Latex
3. Physical Attributes meet or exceed ASTM standards for exam gloves
4. CAUTION: Gloves used for protection against fentanyl and chemotherapy drug exposure must be selected specifically for the type of chemicals used. Review material safety data sheets for the drug being used to determine the required level of protection. Contact Ventyv for specific testing information.

Ventyv<sup>®</sup>  
Outsmart Infection<sup>®</sup>



sritrangusa.com | ventyv.com | Hello@sritrangusa.com | 1.888.483.6898 |  
+1 (813) 606-4301 | 5820 West Cypress Street, Suite H | Tampa, FL 33607  
Issued: 20210913