

Ventyv[®]

Nitrile PF Exam

PLUS 3.5 Black



Ventyv[®]
Outsmart Infection[®]



sritrangusa.com | ventyv.com | Hello@sritrangusa.com | 1.888.483.6898 |
+1 (813) 606-4301 | 5820 West Cypress Street, Suite H | Tampa, FL 33607
Issued: 20210419

R2

Ventyv®

Nitrile PF Exam

PLUS 3.5 Black



Ventyv® Nitrile Powder-Free Plus 3.5 Black (Stallion) Exam Glove has a polymer coating to aid donning and doffing, coupled with superior barrier protection and is ideal for use in a large variety of clinical settings.

Stallion



| Reorder # | Size | Gloves/ Box ¹ | Boxes/ Case |
|-----------|------|-----------------------------|----------------|
| 10333107 | XS | 200 | 10 |
| 10334107 | S | 200 | 10 |
| 10335107 | M | 200 | 10 |
| 10336107 | L | 200 | 10 |
| 10337107 | XL | 180 | 10 |

PRODUCT ATTRIBUTES

- Black
- Nitrile²
- Powder-free
- Polymer Coated
- Tested for Use with Chemotherapy Drugs⁴
- Tested for Use with Fentanyl⁴
- Fingertip Textured

PHYSICAL ATTRIBUTES³

- Length: 220 mm (XS, S)
230 mm (M, L, XL)
- Width: 70 ± 10 mm (XS)
80 ± 10 mm (S)
95 ± 10 mm (M)
110 ± 10 mm (L)
120 ± 10 mm (XL)

THICKNESS (Single Wall)

- Finger: 0.10 mm / 4 mils
- Palm: 0.07 mm / 3 mils

PHYSICAL PROPERTIES

- Tensile strength before aging: 40 MPa
- Tensile strength after aging: 40 MPa
- Elongation before aging: 541%
- Elongation after aging: 500%



80%

Inner boxes made with 80% recycled material

90%

Outer cartons made with 90% recycled material

1. By weight
2. This product is not made with Natural Rubber Latex
3. Physical Attributes meet or exceed ASTM standards for exam gloves
4. CAUTION: Gloves used for protection against fentanyl and chemotherapy drug exposure must be selected specifically for the type of chemicals used. Review material safety data sheets for the drug being used to determine the required level of protection. Contact Ventyv for specific testing information

Ventyv®
Outsmart Infection®



sritrangusa.com | ventyv.com | Hello@sritrangusa.com | 1.888.483.6898 |
+1 (813) 606-4301 | 5820 West Cypress Street, Suite H | Tampa, FL 33607
Issued: 20210419

R2